GEORGIA DRINKING WATER LABORATORY ANALYST PRACTICE TEST (SAMPLE)

STUDY GUIDE



SAMPLE STUDY GUIDE WITH LIMITED QUESTIONS

VISIT US ONLINE FOR THE FULL VERSION STUDY GUIDE

https://gadrinkingwaterlabanalyst.examzify.com

Copyright © 2026 by Examzify - A Kaluba Technologies Inc. product.

ALL RIGHTS RESERVED.

No part of this book may be reproduced or transferred in any form or by any means, graphic, electronic, or mechanical, including photocopying, recording, web distribution, taping, or by any information storage retrieval system, without the written permission of the author.

Notice: Examzify makes every reasonable effort to obtain accurate, complete, and timely information about this product from reliable sources.

SAMPLE

Table of Contents

Copyright 1

Table of Contents 2

Introduction 3

How to Use This Guide 4

Questions 5

Answers 8

Explanations 10

Next Steps 15

SAMPLE

Introduction

Preparing for a certification exam can feel overwhelming, but with the right tools, it becomes an opportunity to build confidence, sharpen your skills, and move one step closer to your goals. At Examzify, we believe that effective exam preparation isn't just about memorization, it's about understanding the material, identifying knowledge gaps, and building the test-taking strategies that lead to success.

This guide was designed to help you do exactly that.

Whether you're preparing for a licensing exam, professional certification, or entry-level qualification, this book offers structured practice to reinforce key concepts. You'll find a wide range of multiple-choice questions, each followed by clear explanations to help you understand not just the right answer, but why it's correct.

The content in this guide is based on real-world exam objectives and aligned with the types of questions and topics commonly found on official tests. It's ideal for learners who want to:

- Practice answering questions under realistic conditions,
- Improve accuracy and speed,
- Review explanations to strengthen weak areas, and
- Approach the exam with greater confidence.

We recommend using this book not as a stand-alone study tool, but alongside other resources like flashcards, textbooks, or hands-on training. For best results, we recommend working through each question, reflecting on the explanation provided, and revisiting the topics that challenge you most.

Remember: successful test preparation isn't about getting every question right the first time, it's about learning from your mistakes and improving over time. Stay focused, trust the process, and know that every page you turn brings you closer to success.

Let's begin.

SAMPLE

How to Use This Guide

This guide is designed to help you study more effectively and approach your exam with confidence. Whether you're reviewing for the first time or doing a final refresh, here's how to get the most out of your Examzify study guide:

1. Start with a Diagnostic Review

Skim through the questions to get a sense of what you know and what you need to focus on. Your goal is to identify knowledge gaps early.

2. Study in Short, Focused Sessions

Break your study time into manageable blocks (e.g. 30 – 45 minutes). Review a handful of questions, reflect on the explanations.

3. Learn from the Explanations

After answering a question, always read the explanation, even if you got it right. It reinforces key points, corrects misunderstandings, and teaches subtle distinctions between similar answers.

4. Track Your Progress

Use bookmarks or notes (if reading digitally) to mark difficult questions. Revisit these regularly and track improvements over time.

5. Simulate the Real Exam

Once you're comfortable, try taking a full set of questions without pausing. Set a timer and simulate test-day conditions to build confidence and time management skills.

6. Repeat and Review

Don't just study once, repetition builds retention. Re-attempt questions after a few days and revisit explanations to reinforce learning. Pair this guide with other Examzify tools like flashcards, and digital practice tests to strengthen your preparation across formats.

There's no single right way to study, but consistent, thoughtful effort always wins. Use this guide flexibly, adapt the tips above to fit your pace and learning style. You've got this!

SAMPLE

Questions

SAMPLE

1. What range of fluoride does the state recommend that a system maintain?

A. 0.1-0.5 mg/L

B. 0.7-1.3 mg/L

C. 1.5-2.0 mg/L

D. 2.0-3.0 mg/L

2. What are the three common fluoride chemicals used in water treatment?

A. Fluorosilicic Acid, Sodium Fluoride, Sodium Fluorosilicate

B. Calcium Fluoride, Sodium Chloride, Potassium Fluoride

C. Fluorinated Hydrocarbons, Trisodium Phosphate, Boron Fluoride

D. Fluorite, Aluminum Sulfate, Chlorinated Hydrocarbons

3. What criteria indicate that samples have been dried to a stable weight?

A. Less than 5% difference

B. Less than 4% or 0.5mg difference

C. Less than 2% difference

D. Less than 1mg difference

4. What is the primary goal of disinfection in water treatment?

A. To enhance taste

B. To remove sediment

C. To eliminate pathogenic organisms

D. To adjust mineral content

5. What is the action level for lead in drinking water?

A. 0.005 mg/L

B. 0.010 mg/L

C. 0.015 mg/L

D. 0.020 mg/L

6. Why is record retention important in water testing?

A. To keep track of testing equipment

B. To document and verify compliance with regulations

C. To enhance the visibility of trends

D. To share with other laboratories

SAMPLE

7. Which parameter must be closely monitored when performing chlorine testing methods?

A. Temperature

B. pH

C. Conductivity

D. Salinity

8. What do the initials TC stand for in laboratory measurements?

A. To Contain

B. To Check

C. To Change

D. To Clean

9. When measuring for phenolphthalein alkalinity, the sample is titrated to an end-point pH of what?

A. pH 7.0

B. pH 8.3

C. pH 4.5

D. pH 6.0

10. What does it indicate when the Colilert sample fluoresces?

A. Salmonella presence

B. Coliform bacteria

C. E. Coli presence

D. None of the above

SAMPLE

Answers

SAMPLE

1. B
2. A
3. B
4. C
5. C
6. B
7. B
8. A
9. B
10. C

SAMPLE

Explanations

SAMPLE

1. What range of fluoride does the state recommend that a system maintain?

A. 0.1-0.5 mg/L

B. 0.7-1.3 mg/L

C. 1.5-2.0 mg/L

D. 2.0-3.0 mg/L

The recommended range of fluoride for a water system to maintain is 0.7 to 1.3 mg/L. This range aligns with public health guidelines aimed at optimizing dental health. Fluoride is added to drinking water to help reduce the incidence of dental caries (cavities) in the population. In this range, the concentration is sufficient to deliver the health benefits associated with fluoride without posing significant health risks. Levels below 0.7 mg/L may not provide adequate dental protection, while levels above 1.3 mg/L could potentially lead to dental fluorosis, a condition that affects the appearance of teeth. Maintaining fluoride within this recommended range helps ensure that drinking water can support community health effectively.

2. What are the three common fluoride chemicals used in water treatment?

A. Fluorosilicic Acid, Sodium Fluoride, Sodium Fluorosilicate

B. Calcium Fluoride, Sodium Chloride, Potassium Fluoride

C. Fluorinated Hydrocarbons, Trisodium Phosphate, Boron Fluoride

D. Fluorite, Aluminum Sulfate, Chlorinated Hydrocarbons

Fluorosilicic Acid, Sodium Fluoride, and Sodium Fluorosilicate are widely recognized as three common fluoride chemicals utilized in water treatment processes for the purpose of fluoridation. These chemicals are chosen primarily for their effectiveness in preventing dental cavities among populations consuming treated water. Fluorosilicic acid, for instance, is a byproduct derived from the production of phosphate fertilizers and is commonly used due to its solubility in water, making it an effective choice for fluoridation. Sodium fluoride is another well-known fluoride compound that is effective in providing the necessary fluoride ions in drinking water. Finally, sodium fluorosilicate serves a similar purpose, acting as a source of fluoride in treated water. The other options contain compounds that are either not primarily used for fluoride treatment or are not relevant to the context of fluoridation in drinking water. Therefore, the selection of the first set of chemicals aligns with standard practices in water treatment aimed at enhancing dental health through controlled fluoridation.

3. What criteria indicate that samples have been dried to a stable weight?

A. Less than 5% difference

B. Less than 4% or 0.5mg difference

C. Less than 2% difference

D. Less than 1mg difference

The correct choice indicates that a sample is considered to be dried to a stable weight when there is either a difference of less than 4% or a weight difference of less than 0.5mg between two successive measurements. This standard ensures that the drying process has effectively removed moisture and that the sample weight has reached a consistent level, indicating minimal further change. Using this specific threshold allows analysts to ascertain that the drying process is thorough. A smaller percentage of difference (like 2% or less) can sometimes be too stringent for practical laboratory procedures, particularly if the mass of the sample is small. Furthermore, relying solely on a weight difference of 1mg might not provide the necessary assurance of stability when the sample size varies. Thus, the combination of both percentage and absolute weight criteria in choice B provides a reliable indicator of stability in weight during the drying of samples.

SAMPLE

4. What is the primary goal of disinfection in water treatment?

A. To enhance taste

B. To remove sediment

C. To eliminate pathogenic organisms

D. To adjust mineral content

The primary goal of disinfection in water treatment is to eliminate pathogenic organisms. Disinfection processes, such as chlorination, UV treatment, or ozonation, are specifically designed to eradicate harmful microorganisms such as bacteria, viruses, and protozoa that can pose serious health risks to humans. By effectively reducing or destroying these pathogens, disinfection ensures that the water is safe for consumption and minimizes the incidence of waterborne diseases. While enhancing taste, removing sediment, and adjusting mineral content are important aspects of overall water treatment, they serve different functions and are not the primary focus of disinfection. The foremost concern during the disinfection process is to safeguard public health by providing microbiologically safe drinking water.

5. What is the action level for lead in drinking water?

A. 0.005 mg/L

B. 0.010 mg/L

C. 0.015 mg/L

D. 0.020 mg/L

The action level for lead in drinking water is set at 0.015 mg/L. This standard is established by the Environmental Protection Agency (EPA) to protect public health, particularly vulnerable populations such as children and pregnant women, from the harmful effects of lead exposure. If the lead concentration in drinking water exceeds this level, water systems must take specific actions to reduce lead levels, including notifying the public and implementing corrosion control measures. This action level is critical for ensuring safe drinking water and aligns with regulatory requirements designed to minimize lead contamination in water supplies.

6. Why is record retention important in water testing?

A. To keep track of testing equipment

B. To document and verify compliance with regulations

C. To enhance the visibility of trends

D. To share with other laboratories

Record retention is crucial in water testing primarily to document and verify compliance with regulations. Water quality testing must adhere to specific state and federal standards, and maintaining accurate records provides evidence that these tests have been conducted properly and in accordance with the law. This documentation is essential during audits, inspections, and assessments, ensuring that the laboratory is operating within legal requirements and that water supplies are safe for public consumption. Additionally, proper record retention aids in accountability, allowing any discrepancies or issues to be traced back through historical data. This is vital for governmental bodies and regulatory agencies in enforcing laws designed to protect public health. When records are meticulously kept, it builds trust with stakeholders, including the community, regulatory agencies, and other entities involved in water management and public health.

SAMPLE

7. Which parameter must be closely monitored when performing chlorine testing methods?

A. Temperature

B. pH

C. Conductivity

D. Salinity

The parameter that must be closely monitored when performing chlorine testing methods is pH. Chlorine's disinfecting ability is significantly influenced by the pH level of the water. At different pH levels, chlorine exists in different forms: free chlorine (hypochlorous acid and hypochlorite ion) and combined chlorine. Hypochlorous acid is the more effective form of chlorine for disinfection, and its concentration is highest at a pH level around 7.2 to 7.6. If the pH is too high or too low, it can alter the effectiveness of chlorine as a disinfectant, leading to insufficient disinfection or the formation of undesirable chlorination byproducts. In contrast, while temperature, conductivity, and salinity can influence various chemical processes in water, they are not as critical to chlorine effectiveness as pH. Temperature can affect the rate of chlorination reactions but does not change the chlorine's effectiveness in the same specified way. Conductivity and salinity are more relevant to other aspects of water quality and treatment rather than specifically monitoring chlorine efficacy. Therefore, keeping a close watch on pH is essential for ensuring optimal chlorine performance in water treatment processes.

8. What do the initials TC stand for in laboratory measurements?

A. To Contain

B. To Check

C. To Change

D. To Clean

In laboratory measurements, the initials TC stand for "To Contain." This designation indicates that a container or measuring device is calibrated to hold a specific volume of liquid, but does not account for the liquid that may remain on the inside walls of the container when it is poured out. This is particularly important in laboratory settings where precise volume measurements are crucial for accurate testing and analyses. Understanding the meaning of TC is vital for laboratory analysts, as it affects how one interprets the results of volumetric measurements. The term emphasizes the importance of proper dispensing techniques to ensure the accuracy of liquid volumes used in experiments. This knowledge helps in maintaining consistency and reliability in test results, which is fundamental in drinking water analysis and compliance with regulatory standards.

SAMPLE

9. When measuring for phenolphthalein alkalinity, the sample is titrated to an end-point pH of what?

A. pH 7.0

B. pH 8.3

C. pH 4.5

D. pH 6.0

The correct approach for measuring phenolphthalein alkalinity involves titrating the water sample to an endpoint pH of 8.3. At this pH, phenolphthalein changes color, which indicates the transition point between acid and base conditions in relation to the presence of bicarbonates and carbonates in the sample. Phenolphthalein alkalinity is primarily associated with the presence of carbonate ions, which contributes to the buffering capacity of water. A pH of 8.3 aligns with the typical transition point where carbonate ions begin to deprotonate to form bicarbonate ions. Thus, reaching this specific pH during titration signifies that carbonate is buffering the pH up to this point, effectively measuring the strength of the alkalinity attributable to this component. Understanding the specific pH endpoint is essential for accurately assessing water quality, particularly in evaluating its capability to neutralize acids and maintain pH stability. This knowledge is crucial for laboratory analysts working with drinking water to ensure it meets health and safety standards.

10. What does it indicate when the Colilert sample fluoresces?

A. Salmonella presence

B. Coliform bacteria

C. E. Coli presence

D. None of the above

When a Colilert sample fluoresces, it is an indication of the presence of E. coli. The Colilert test uses a medium that contains nutrients and indicator compounds that specifically react with coliform bacteria. In the presence of E. coli, the enzymes produced by these bacteria act on the compound in the reagent, resulting in fluorescence. This is a significant aspect of using the Colilert test, as it allows for the rapid and specific identification of E. coli, which is a crucial marker for fecal contamination in drinking water. The other choices do not reflect the specific outcome of fluorescence in the Colilert test. While coliform bacteria are indeed detected in the test, they do not cause fluorescence; the test is designed specifically to highlight the presence of E. coli through this reaction. Salmonella, on the other hand, is not detected by the Colilert method, which focuses primarily on coliforms and E. coli. Therefore, the accurate identification of E. coli through fluorescence underlines its importance in evaluating water safety and quality.

SAMPLE

Next Steps

Congratulations on reaching the final section of this guide. You've taken a meaningful step toward passing your certification exam and advancing your career.

As you continue preparing, remember that consistent practice, review, and self-reflection are key to success. Make time to revisit difficult topics, simulate exam conditions, and track your progress along the way.

If you need help, have suggestions, or want to share feedback, we'd love to hear from you. Reach out to our team at hello@examzify.com.

Or visit your dedicated course page for more study tools and resources:

https://gadrinkingwaterlabanalyst.examzify.com

We wish you the very best on your exam journey. You've got this!

SAMPLE