POSITRON EMISSION TOMOGRAPHY–COMPUTED TOMOGRAPHY (PET/CT) FUSION PRACTICE EXAM (SAMPLE)

STUDY GUIDE



SAMPLE STUDY GUIDE WITH LIMITED QUESTIONS

VISIT US ONLINE FOR THE FULL VERSION STUDY GUIDE

https://petctfusion.examzify.com

Copyright © 2026 by Examzify - A Kaluba Technologies Inc. product.

ALL RIGHTS RESERVED.

No part of this book may be reproduced or transferred in any form or by any means, graphic, electronic, or mechanical, including photocopying, recording, web distribution, taping, or by any information storage retrieval system, without the written permission of the author.

Notice: Examzify makes every reasonable effort to obtain accurate, complete, and timely information about this product from reliable sources.

SAMPLE

Table of Contents

Copyright 1

Table of Contents 2

Introduction 3

How to Use This Guide 4

Questions 5

Answers 8

Explanations 10

Next Steps 16

SAMPLE

Introduction

Preparing for a certification exam can feel overwhelming, but with the right tools, it becomes an opportunity to build confidence, sharpen your skills, and move one step closer to your goals. At Examzify, we believe that effective exam preparation isn't just about memorization, it's about understanding the material, identifying knowledge gaps, and building the test-taking strategies that lead to success.

This guide was designed to help you do exactly that.

Whether you're preparing for a licensing exam, professional certification, or entry-level qualification, this book offers structured practice to reinforce key concepts. You'll find a wide range of multiple-choice questions, each followed by clear explanations to help you understand not just the right answer, but why it's correct.

The content in this guide is based on real-world exam objectives and aligned with the types of questions and topics commonly found on official tests. It's ideal for learners who want to:

- Practice answering questions under realistic conditions,
- Improve accuracy and speed,
- Review explanations to strengthen weak areas, and
- Approach the exam with greater confidence.

We recommend using this book not as a stand-alone study tool, but alongside other resources like flashcards, textbooks, or hands-on training. For best results, we recommend working through each question, reflecting on the explanation provided, and revisiting the topics that challenge you most.

Remember: successful test preparation isn't about getting every question right the first time, it's about learning from your mistakes and improving over time. Stay focused, trust the process, and know that every page you turn brings you closer to success.

Let's begin.

SAMPLE

How to Use This Guide

This guide is designed to help you study more effectively and approach your exam with confidence. Whether you're reviewing for the first time or doing a final refresh, here's how to get the most out of your Examzify study guide:

1. Start with a Diagnostic Review

Skim through the questions to get a sense of what you know and what you need to focus on. Your goal is to identify knowledge gaps early.

2. Study in Short, Focused Sessions

Break your study time into manageable blocks (e.g. 30 – 45 minutes). Review a handful of questions, reflect on the explanations.

3. Learn from the Explanations

After answering a question, always read the explanation, even if you got it right. It reinforces key points, corrects misunderstandings, and teaches subtle distinctions between similar answers.

4. Track Your Progress

Use bookmarks or notes (if reading digitally) to mark difficult questions. Revisit these regularly and track improvements over time.

5. Simulate the Real Exam

Once you're comfortable, try taking a full set of questions without pausing. Set a timer and simulate test-day conditions to build confidence and time management skills.

6. Repeat and Review

Don't just study once, repetition builds retention. Re-attempt questions after a few days and revisit explanations to reinforce learning. Pair this guide with other Examzify tools like flashcards, and digital practice tests to strengthen your preparation across formats.

There's no single right way to study, but consistent, thoughtful effort always wins. Use this guide flexibly, adapt the tips above to fit your pace and learning style. You've got this!

SAMPLE

Questions

SAMPLE

1. Which physiological processes can be evaluated using PET imaging?

A. Bone density and heart rhythm

B. Glucose metabolism and blood flow

C. Visual acuity and auditory function

D. Respiratory function and muscle strength

2. What is the typical duration of a PET/CT scan appointment?

A. Approximately 1 hour

B. Approximately 2 hours, including preparation and imaging time

C. Approximately 30 minutes

D. Approximately 3 hours

3. What should be a primary consideration during the patient consent process for PET/CT scans?

A. Length of the scan procedure.

B. Understanding potential risks and benefits.

C. Personal medical history only.

D. Cost of the procedure.

4. What is one of the main reasons for performing attenuation correction in PET imaging?

A. To enhance patient comfort

B. To improve image resolution

C. To provide accurate SUV calculations

D. To reduce the scan time

5. How does density of the surrounding tissue affect PET imaging?

A. It influences the risk of false negatives

B. It determines the choice of imaging frequency

C. It impacts attenuation and resolution of the images

D. It alters the positron emission rate

SAMPLE

6. What is the purpose of quality control in imaging procedures?

A. To ensure consistent and accurate results across different scans

B. To minimize the dose of radiation to patients

C. To reduce the duration of the imaging process

D. To improve the comfort of patients during scans

7. What can cause a false positive result in a PET scan?

A. Recent surgical procedures

B. Inflammation or infection, which may also take up FDG

C. Improper patient positioning

D. Low-quality imaging equipment

8. What should be done to correct for random events in PET imaging?

A. Simultaneously acquire in 2 energy windows

B. Create a map of density within the field of view

C. Inject the proper dose

D. Increase image time

9. How is radiation exposure in PET/CT imaging generally compared to traditional CT scans?

A. It has the same level of exposure

B. It is generally lower than traditional CT scans

C. It is typically higher due to the radiotracer

D. It varies significantly based on the patient

10. What is one of the primary advantages of using PET/CT in clinical practice?

A. Decreased cost of treatment

B. Increased detail and exact localization of tumors

C. Reduction of scan duration

D. Elimination of radiation exposure

SAMPLE

Answers

SAMPLE

1. B
2. B
3. B
4. C
5. C
6. A
7. B
8. C
9. C
10. B

SAMPLE

Explanations

SAMPLE

1. Which physiological processes can be evaluated using PET imaging?

A. Bone density and heart rhythm

B. Glucose metabolism and blood flow

C. Visual acuity and auditory function

D. Respiratory function and muscle strength

PET imaging is particularly useful in evaluating various physiological processes, with its most prominent application being the assessment of metabolic activity. One of the primary aspects that can be examined through PET is glucose metabolism. This is because PET scans often utilize radiolabeled glucose analogs, such as fluorodeoxyglucose (FDG), which is taken up by metabolically active tissues, making it an excellent tool for identifying areas of increased metabolic activity, such as tumors. Additionally, PET imaging is effective in evaluating blood flow. This function is significant in the assessment of various conditions, including cardiovascular diseases and cancer, as it provides insight into how blood circulates through tissues, which can indicate perfusion abnormalities. In contrast, other options such as bone density and heart rhythm, visual acuity and auditory function, and respiratory function and muscle strength, are outside the primary scope of PET imaging. These processes are better assessed with other imaging modalities or tests that are specifically designed for those purposes.

2. What is the typical duration of a PET/CT scan appointment?

A. Approximately 1 hour

B. Approximately 2 hours, including preparation and imaging time

C. Approximately 30 minutes

D. Approximately 3 hours

The typical duration of a PET/CT scan appointment includes various phases such as preparation, the scanning process itself, and recovery time. The correct answer reflects the comprehensive nature of the procedure, which can take about 2 hours in total. During this time, the patient goes through several important steps. Initially, there is preparation time where the patient may need to fast, depending on the type of scan, allowing the radiotracer to be taken up effectively. Following preparation, the actual imaging part involves the injection of the radioactive tracer, followed by a waiting period of about 30 to 60 minutes to allow for adequate uptake before the CT scan. The imaging itself usually lasts between 20 to 30 minutes, but additional time for positioning, setup, and potential post-scan observations contributes to the overall duration. Hence, accounting for all these factors, a total appointment duration of approximately 2 hours is typical for a PET/CT scan, making this answer the most accurate.

SAMPLE

3. What should be a primary consideration during the patient consent process for PET/CT scans?

A. Length of the scan procedure.

B. Understanding potential risks and benefits.

C. Personal medical history only.

D. Cost of the procedure.

The primary consideration during the patient consent process for PET/CT scans is understanding potential risks and benefits. This is vital because informed consent involves ensuring that patients are fully aware of what the procedure entails, its necessity, and any possible adverse effects that may arise from it. PET/CT scans involve the use of a radioactive tracer, and it is important for patients to comprehend both the diagnostic value of the scan and the implications of radiation exposure. By clearly discussing the potential risks—such as allergic reactions to the tracer or the cumulative effect of radiation with previous procedures—as well as the benefits, like accurately identifying conditions such as cancer or assessing treatment response, healthcare providers enable patients to make informed decisions regarding their healthcare. This comprehensive understanding empowers patients and helps them feel more involved in their medical care. While aspects such as the length of the scan procedure, personal medical history, and the cost of the procedure are important considerations, they do not hold the same weight in the context of informed consent as understanding the risks and benefits of the PET/CT scan does.

4. What is one of the main reasons for performing attenuation correction in PET imaging?

A. To enhance patient comfort

B. To improve image resolution

C. To provide accurate SUV calculations

D. To reduce the scan time

The primary purpose of performing attenuation correction in PET imaging is to provide accurate Standardized Uptake Value (SUV) calculations. Attenuation refers to the reduction in the intensity of the radiation emitted from positron-emitting tracers due to absorption and scattering by body tissues. If this attenuation is not properly accounted for, it can lead to inaccurate measurements of radioactivity within the tissues, ultimately affecting the SUV values. Accurate SUV calculations are essential for evaluating metabolic activity in tumors and other pathological processes, allowing for better diagnosis and treatment planning. This correction process helps ensure that the measured tracer uptake reflects true biological activity rather than artifacts caused by tissue density variations. In clinical practice, precise SUV values enable comparisons over time and between different patients, making this correction crucial for consistent and reliable interpretation of PET scans.

SAMPLE

5. How does density of the surrounding tissue affect PET imaging?

A. It influences the risk of false negatives

B. It determines the choice of imaging frequency

C. It impacts attenuation and resolution of the images

D. It alters the positron emission rate

The density of the surrounding tissue plays a crucial role in PET imaging by impacting attenuation and resolution of the images. In PET scans, positrons emitted by radiopharmaceuticals collide with electrons in the surrounding tissues, resulting in annihilation and the emission of gamma rays. The density of the tissue directly affects how well these gamma rays can penetrate through the body. Higher-density tissues, such as bone, attenuate the gamma rays more than lower-density tissues, such as fat or soft tissue. This attenuation can lead to a reduction in the number of detected photons reaching the PET scanner, which ultimately impacts the quality and resolution of the images produced. If the attenuation is significant, it can obscure the visibility of tumors or other abnormalities, leading to potential misinterpretations of the scans. Understanding how different tissue densities influence these factors is essential for interpreting PET scans accurately, as it can help technologists and radiologists make informed decisions about the imaging process, including potential adjustments to improve image quality.

6. What is the purpose of quality control in imaging procedures?

A. To ensure consistent and accurate results across different scans

B. To minimize the dose of radiation to patients

C. To reduce the duration of the imaging process

D. To improve the comfort of patients during scans

Quality control in imaging procedures is essential for ensuring consistent and accurate results across different scans. It involves a series of systematic measures that verify and calibrate the imaging equipment, evaluate the imaging process, and assess the quality of the images produced. This is crucial because variations in equipment performance, operator technique, or environmental factors can lead to discrepancies in imaging results. Accurate imaging is vital for diagnosis, treatment planning, and monitoring of diseases. Without routine quality control, there is a risk of obtaining misleading images that could affect patient management. Thus, implementing comprehensive quality control protocols helps maintain the reliability of the imaging results, ensuring that healthcare providers can make informed decisions based on high-quality data. While minimizing radiation exposure, reducing imaging duration, and enhancing patient comfort are important considerations in the imaging process, the primary purpose of quality control is to guarantee the integrity and reliability of the imaging results. Therefore, focusing on achieving consistent and accurate outcomes is the cornerstone of quality control in imaging procedures.

SAMPLE

7. What can cause a false positive result in a PET scan?

A. Recent surgical procedures

B. Inflammation or infection, which may also take up FDG

C. Improper patient positioning

D. Low-quality imaging equipment

A false positive result in a PET scan can occur when areas of increased glucose metabolism, detected by the radioactive tracer FDG (Fluorodeoxyglucose), are present due to conditions other than cancer. Inflammation or infection can significantly elevate FDG uptake because these processes often involve increased cellular activity, leading to a higher rate of glucose metabolism in affected tissues. For example, conditions such as pneumonia, abscesses, or even postoperative inflammation can result in areas of the body showing increased FDG uptake, mimicking the signatures typically associated with tumor presence. This is a particularly important consideration in interpreting PET scan results, as distinguishing between these benign processes and malignancy can be challenging. Other potential factors influencing test results include surgical procedures, which may also lead to localized inflammatory responses; however, inflammation or infection are more direct causes of false positives due to their nature of increasing metabolic activity that the PET scan captures. Proper patient positioning and high-quality imaging equipment primarily pertain to the technical execution of the scan rather than pathological reasons for false positives.

8. What should be done to correct for random events in PET imaging?

A. Simultaneously acquire in 2 energy windows

B. Create a map of density within the field of view

C. Inject the proper dose

D. Increase image time

In the context of correcting for random events in PET imaging, the emphasis is primarily on ensuring optimal conditions for capturing accurate signals from tracer distribution. The correct choice involves injecting the proper dose of the radiotracer. Administering an appropriate amount of the radiotracer is crucial for distinguishing the signal from random coincidences, which are events that occur due to detector noise rather than actual positron emissions. When a sufficient dose is administered, it increases the likelihood of capturing genuine coincidences between detected photons from the annihilation events, thereby enhancing the signal-to-noise ratio. This helps in minimizing the impact of random events during the imaging process, leading to improved image quality and accuracy in quantifying radioactive uptake in tissues. Meanwhile, the other options, while they may seem applicable in different contexts, do not directly address the correction of random events related to tracer administration in the same effective manner as ensuring the correct dose does. Simultaneously acquiring in two energy windows is often used in advanced techniques but is not primarily aimed at correcting random events. Creating a density map is for different diagnostic purposes, and simply increasing the image time does not resolve the underlying issues related to random coincidences but rather may contribute to longer acquisition times without ensuring signal clarity.

SAMPLE

9. How is radiation exposure in PET/CT imaging generally compared to traditional CT scans?

A. It has the same level of exposure

B. It is generally lower than traditional CT scans

C. It is typically higher due to the radiotracer

D. It varies significantly based on the patient

In PET/CT imaging, radiation exposure is typically higher compared to traditional CT scans due to the additional radiotracer injected into the patient. The PET component of the scan involves the use of a radioactive substance, which is necessary for the imaging process as it allows visualization of metabolic processes in the body. This radiotracer emits positrons, which interact with electrons in the body and produce gamma rays that are detected by the PET scanner. The amount of radiation from the CT scan itself can vary, but when combined with the radiation from the radiotracer used in PET imaging, the total exposure tends to exceed that of a standalone traditional CT scan. While traditional CT scans do involve exposure to ionizing radiation, the additional dose from the PET component elevates the overall exposure in PET/CT imaging. Understanding this aspect is crucial for healthcare providers when considering the risks and benefits of using PET/CT for diagnosis and treatment planning.

10. What is one of the primary advantages of using PET/CT in clinical practice?

A. Decreased cost of treatment

B. Increased detail and exact localization of tumors

C. Reduction of scan duration

D. Elimination of radiation exposure

The primary advantage of using PET/CT in clinical practice lies in its ability to provide increased detail and exact localization of tumors. PET (Positron Emission Tomography) is excellent for visualizing metabolic activity in tissues, while CT (Computed Tomography) provides detailed anatomical information. When these two imaging modalities are combined, they allow healthcare professionals to pinpoint the location of tumors with a high level of accuracy and understand their metabolic characteristics. This is particularly beneficial for cancer diagnosis, staging, and treatment planning, as it enhances the ability to differentiate between benign and malignant lesions and assess treatment response. The synergy of functional and anatomical imaging that PET/CT offers is a significant factor in its effectiveness in clinical applications. While factors like cost reduction, scan duration, and minimizing radiation exposure are important considerations in medical imaging, they do not represent the core advantage of PET/CT technology compared to traditional imaging methods. The emphasis on detailed tumor localization fundamentally improves diagnostic and therapeutic outcomes.

SAMPLE

Next Steps

Congratulations on reaching the final section of this guide. You've taken a meaningful step toward passing your certification exam and advancing your career.

As you continue preparing, remember that consistent practice, review, and self-reflection are key to success. Make time to revisit difficult topics, simulate exam conditions, and track your progress along the way.

If you need help, have suggestions, or want to share feedback, we'd love to hear from you. Reach out to our team at hello@examzify.com.

Or visit your dedicated course page for more study tools and resources:

https://petctfusion.examzify.com

We wish you the very best on your exam journey. You've got this!

SAMPLE